

Power Meter Wavelength Dependence

Typical Power Meter Uncertainty (dB) due to ± 30 nm wavelength tolerance

	Si	Ge	InGaAs
850 nm	0.10	0.46	1.30 dB
1300/1300 nm		0.06	0.01 dB
1550 nm		1.00	0.05 dB

