



DECLARATION OF CONFORMITY
IN ACCORDANCE WITH ISO/IEC 17050:2004



Manufacturer's Name: Kingfisher International Pty. Ltd.
Manufacturer's Address: 30 Rocco Drive, Scoresby, Victoria 3179, Australia

hereby declares, that the products listed below

Product Name: Light Source, Power Meter
Model Number: KI9800 Series, KI9600 Series
Product Options: This declaration covers all options of the above product(s)

comply with the essential requirements of the applicable European Directives:

- Low Voltage Directive 73/23/EEC and the EMC Directive 89/336/EEC, amended by 93/68/EEC, and carries the CE marking accordingly
- Directive 2002/95/EC on restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)
- Directive 2002/96/EC on waste electrical and electronic equipment (WEEE)

and conform to the following standards and specifications:

MIL-PRF-28800F: 1996 Performance specification-Test equipment for use with electrical and electronic equipment, general specification
IEC 60529: 2001/ EN 60529:1993+A1:2003 Degrees of protection provided by enclosures

Power Meter conforms to:
IEC 61315: 2005 Calibration of fibre-optic power meters

EMC

IEC 61326:2002 / EN 61326-1:1997+A1:1998+A2:2001+A3:2003
IEC CISPR 11: 2004/ EN 55011:1998+A1:1999
IEC CISPR 16-1:1999
IEC CISPR 16-2:1999
IEC 61000-4-2:2001 / EN 61000-4-2:1995+A1:1998+A2:2001
IEC 61000-4-3:2002/ EN 61000-4-3:2002
ICES-001: 2006 (Canada)
CFR 47 FCC Part 15, Subpart B (Class B) (USA)
FCC registration number: 90891

Limit

Limits applicable to Group 1 Class B equipment
Limits applicable to Class B equipment
Limits applicable to Class B equipment
4kV CD, 8kV AD
3V/m, 80-1000MHz


Safety

Light Source conforms to:
IEC 60825-1:2001 Safety of laser products - Equipment classification, requirements and user's guide
IEC 60825-2:2005 Safety of laser products - Safety of optical fibre communication systems (OFCS)
CFR 21 part 1040.10 (USA) Performance standards for light- emitting products-Laser products

Supplemental Information:

The product was tested in a typical configuration with Kingfisher International test systems.

2008-December17
Date


Bruce Robertson
Name

Technical Director
Title

For further information, please contact your local Kingfisher International sales office, agent or distributor.